



Analysis of the Major Elements in Cement by ICP

Nathalie Le Corre
Jobin Yvon S.A.S., Horiba Group
Longjumeau, France

Keywords: cement

1 Introduction

Chemical analysis of cement is an interesting application for ICP spectrometry. The cement industry requires analysis of major elements at % levels to very high accuracy and precision, including Si, Ca, Al, Fe and S and also minor elements such as Na, K, Ti, Mg, P, Mn. The trace elements are also very important for environmental pollution control, this includes species like Cr, Pb, Zn, Ni, As, Cd, V and Cu at ppm levels. All these elements are possible by the ICP technique. The use of a combined ICP-OES, such as the ULTIMA 2C, enables fast analysis of all above mentioned elements. For major elements, the repeatability required is in the range 0.2 to 0.5 %, for this reason the simultaneous component is often used for routine analysis, whereas trace elements are better done on the sequential portion for the superior detection limits.

2 Principle

2.1 Technique used

The elemental analysis of solutions was undertaken by Inductively Coupled Plasma Optical Emission Spectrometry (ICP-OES). The sample is nebulized then transferred to an argon plasma. It is decomposed, atomized and ionized whereby the atoms and ions are excited. We measure the intensity of the light emitted when the atoms or ions return to lower levels of energy. Each element emits light at characteristic wavelengths and these lines can be used for quantitative analysis after a calibration.

2.2 Wavelength choice

The choice of the wavelength in a given matrix can be made using the "profile" function, or by using Win-IMAGE, which is rapid semi-quantitative analysis mode using multiple wavelengths. The principle is the same in either case: record the scans of analytes at low concentration, and of the matrix. By superimposing the spectra, we see possible interferences.

3 Sample Preparation

3.1 Classical digestion

0.1g of sample is fused at 1,000 °C with 1 g of lithium tetraborate for about 30 minutes. The fusion residue is then dissolved with 5 % HCl and made up to a volume of 500 mL.

3.2 Microwave digestion

0.1 g of cement is weighed and 2.5 mL of concentrated HNO₃, 2.5 ml HBF₄ and 0.5 ml concentrated HCL are added to the sample. The digestion program is 5 minutes at 25 % power and then 5 minutes at 50 % power. The residue is dissolved in 5 % HCl and brought up to a volume of 500 mL.



4. Instrument specification

The work was done on a JY ULTIMA 2C. The specifications of this instrument are listed below in Tables 1 and 2.

Table 1: Specification of spectrometer

Parameters	Monochromator	Polychromator
Mounting	Czerny-Turner	Paschen-Runge
Focal length	1 m	0.5 m
Grating number of grooves	2 400 gr/mm	3 600 gr/mm
	working in 2 orders	
Variable resolution	Yes	No
Nitrogen purge	Yes	Yes

Table 2: Specification of RF Generator

Parameters	Specifications
Observation	Radial
Frequency	40.68 MHz
Control of gas flowrate	By computer
Control of pump flow	By computer
Cooling	Air

5 Operating conditions

The operating conditions are listed in Table 3 below.

Table 3: Operating conditions

Parameter	Condition
RF Generator power	1200 W
Plasma gas flowrate	14 L/min
Auxiliary gas flowrate	0.8 L/min
Sheath gas flowrate	0.3 L/min
Nebulizer flowrate	2.51 bars
Sample uptake	0.8 mL/min
Type of nebulizer	JY 0.7
Type of spray chamber	Scott
Argon humidifier	No
Injector tube diameter	3.0 mm

6 Wavelength selection and analytical conditions

All the lines were measured with the direct peaking mode. The integration time was 2 s for the elements analyzed with the monochromator, and 10 s for the polychromator elements.

Table 4: Wavelength selection and analytical conditions

Element	Wavelength
Al	308.215
As	189.042
Ca	317.215
Cd	228.802
Cr	267.716
Fe	259.940
Mg	279.079
Mn	257.610
Na	589.592
Ni	231.604
Pb	220.353
S	180.676
Si	251.611
Tl	190.864
Ti	337.280
V	292.402
Zn	213.856

7 Results

7.1 Standard preparation

The standards were prepared like the samples. For major elements, it is suggested to take at least 4 certified materials. For trace elements, 2 standards are enough. The choice of the standards will depend on the concentration ranges of the samples. For Ca, Si, Fe and Al an internal standard (Li) is required to compensate for the matrix differences.

7.2 Analysis of 3 cements on the monochromator

The cements were measured three times and reported in the table below, with the concentrations expressed in %. The results are presented in Table 5.

7.3 Analysis of 3 cements on the polychromator

The accuracy with the polychromator was as good as for the monochromator, but the precision and repeatability was much better. The results are presented in Table 6.

**Table 5: Analysis of 3 cements on the monochromator**

Element	Conc %	8889 RSD %	Certified %	Conc %	9090 RSD %	Certified %	Conc %	9191 RSD %	Certified %
SO ₃	2.76	1.4	2.77	1.97	0.78	2.03	3.15	1.1	3.07
SiO ₂	25.35	0.17	25.46	21.42	0.16	21.51	21.20	0.26	21.10
Fe ₂ O ₃	2.82	0.33	3.02	3.79	0.19	3.95	4.05	0.48	4.19
MgO	0.705	0.23	0.73	1.02	0.13	1.04	0.694	0.32	0.71
Al ₂ O ₃	3.42	0.46	3.38	4.15	0.28	4.13	3.28	0.18	3.30
CaO	64.60	0.19	65.29	64.30	0.23	65.29	62.13	0.31	61.99
Ti	0.155	0.36	-	0.195	0.51	-	0.18	0.25	-
Na ₂ O	0.088	0.51	0.09	0.17	0.42	0.13	0.134	0.70	0.12
K ₂ O	0.190	0.32	0.19	0.60	1.04	0.59	0.864	0.49	0.82

Table 6: Analysis of 3 cements on the polychromator

Element	Conc %	8889 RSD %	Certified %	Conc %	9090 RSD %	Certified %	Conc %	9191 RSD %	Certified %
SO ₃	2.79	0.31	2.77	2.11	0.17	2.03	3.15	0.23	3.07
SiO ₂	25.38	0.02	25.46	21.49	0.11	21.51	3.96	0.08	21.10
Fe ₂ O ₃	2.78	0.09	3.02	3.86	0.04	3.95	21.27	0.07	4.19
MgO	0.69	0.17	0.73	1.02	0.07	1.04	0.702	0.24	0.71
Al ₂ O ₃	3.35	0.10	3.38	4.14	0.05	4.13	3.29	0.12	3.30
CaO	64.33	0.06	65.29	65.11	0.12	65.29	62.08	0.22	61.99
Ti	0.160	0.14	-	0.202	0.07	-	0.18	0.31	-
Na ₂ O	0.092	1.1	0.09	0.16	0.36	0.13	0.131	0.61	0.12
K ₂ O	0.188	0.26	0.19	0.57	0.18	0.59	0.857	0.34	0.82

8 Conclusion

It is particularly interesting to point out that JY ICPs have the ability to analyze any type of sample for example, cements, where there is a high salt content due to alkali fusion. The advantage of the JY ULTIMA 2C is to simultaneously measure the major and trace elements with very high precision. It also enables you to measure an internal standard at the

same time and with the same optics for all major elements, improving the repeatability and accuracy for these elements. The use of the sheath gas device originally patented by JY, enables analysis without clogging of the injector tube and to be able to run samples up to 300 g/L salt content. Given the above criteria, ICP emission spectrometry is suitable for analysis of the major and minor elements in cement.



In the USA:
Jobin Yvon Inc.
3880 Park Avenue
Edison, NJ 08820
Tel: 1-732-494-8660
Fax: 1-732-494-8796
E-mail:
emission@jyhoriba.com

In France:
Jobin Yvon S.A.S.
16-18, rue du Canal
91165 Longjumeau Cedex
Tel: (33) 1/64 54 13 00
Fax: (33) 1/69 09 90 88

In Japan:
Horiba Ltd.
2 Miyano Higashi, Kisshoin
Minami-ku, Kyoto 601-8510
TEL: (81) 75 313 8121
FAX: (81) 75 321 5725
www.horiba.com

Germany: (49) 89/46 23 17-0
Italy: (39) 2/57 60 56 90
U.K.: (44) 20/82 04 81 42

China: (86) 10/6836 6542
Spain: (34) 91/724 16 57
Other Countries: Contact JY S.A.S.

1-877-JYHORIBA

www.jyhoriba.com



HORIBAGROUP

EMISSION • FLUORESCENCE • FORENSICS • GRATINGS & OEM • RAMAN • OPTICAL SPECTROSCOPY • THIN FILM